

Low Level Laser Therapy for Dentinal Hypersensitivity - A Review

Dr. Kamal Garg,¹ Dr. Rajni Aggarwal,² Dr. Piyush Pareek³

1. Dr. Kamal Garg

professor and HOD, Surendera Dental College and Research Institute, Sri Ganga Nagar. Rajasthan

2. Dr. Rajni Aggarwal .

Reader , Surendera Dental College and Research Institute, Sri Ganga Nagar. Rajasthan

3. Dr. Piyush Pareek

Post Graduate Student. Surendera Dental College and Research Institute, Sri Ganga Nagar. Rajasthan

Abstract Dentinal hypersensitivity (DH) is characterized by a nonspontaneous, acute short- or long-lasting pain originating from exposure of the dentinal tubules or dentine to the thermal, chemical, mechanical, or osmotic stimuli, which cannot be ascribed to any other dental defect or pathology. There are various treatment modalities, one amongst them is lasers. Lasers have been shown to fulfill the requirements of Grossman's criteria by being nonirritating to the pulp. Different lasers have been used for the treatment of dentine hypersensitivity i.e, Low output power (low-level) lasers like (HeNe) helium-neon and (GaAlAs) gallium-aluminumarsenide (diode) lasers and high output laser i.e, (Carbon Dioxide Laser (CO₂), neodymium- or erbium-doped yttriumaluminum garnet (Nd:YAG, Er:YAG lasers) and erbium, chromium doped: yttrium, scandium, gallium and garnet (Er,Cr:YSGG) lasers). Each laser has different mechanism of action in causing a reduction in dentinal hypersensitivity. In case of low level lasers, a small fraction of the lasers energy is transmitted through enamel or dentin to reach the pulp tissue. Low-power laser therapy is an appropriate treatment strategy to promote biomodulatory effects, minimize pain and reduce inflammatory processes. The low-level lasers produce their effects from photobiostimulation effect within the tissues and do not cause temperature elevation within the tissues. The output powers range from 50 to 500 mW with wavelengths in the red and near infrared of the electromagnetic spectrum, from 630 to 980 nm with pulsed or continuous-wave emission. The application of LLLT has become popular in the clinical application for dentinal hypersensitivity. The aim of this paper is to review low-level laser therapy in dentin hypersensitivity.

Keywords: Dentinal hypersensitivity, low level laser therapy.

INTRODUCTION

Dentine hypersensitivity associated with the neck of the tooth is the enigma amongst all the odontological problems associated with pain which has a difficult solution for the dentist. Dentine Hypersensitivity (DH) is an abnormal response of the exposed vital dentine to thermal, chemical, or tactile stimuli, and which cannot be ascribed to any other form of dental defect or pathology.¹ DH affects one in every seven adult patients, affecting

both the sexes with no significant difference and the most affected age group is 20–30 years old.² Cervical areas of buccal surface of canines and premolars have a larger frequency of occurrence of this disorder than other teeth.³ The occurrence of DH is due to the denudation or alteration of the outer layer of the tooth i.e enamel and / or root cementum by processes that include attrition, abrasion and erosion, or by radicular denudation as

a result of gingival recession, any periodontal disease or any post periodontal surgery.⁴

Although different theories have been proposed for the mechanism involved in DH etiology like neural theory, odontoblastic theory but recent studies gave support to Brannstrom's hydrodynamic theory, which states that, a stimulus applied to open dentinal tubules, increases the flow of dentinal tubular fluid, with mechanical deformation of the nerves located into the inner ends of the tubules or in the outer layers of the pulp. Type A delta fibers are supposed to be responsible for dentinal sensitivity being probably activated by the hydrodynamic process.⁵⁻⁶

In order to eliminate such discomfort for patients, different treatment modalities having several substances like desensitizing pastes, varnishes, sealants (i.e for both home and office technique) iontophoresis, fillings have been tested. Most of the treatment modalities have aimed to block exposed dentinal tubules, but none of these treatments has produced consistently effective or long-lasting results without any black points. Seeking a solution for this problem, keeping in mind all the requirements and black points of the previously used treatment solutions or agents and also considering the technological development, the laser therapy was introduced as an alternative for the management of dentin hypersensitivity.

Laser (acronym for LIGHT AMPLIFICATION BY STIMULATED EMISSION OF RADIATION) therapy was first introduced as a potential method for treating DH in the mid-1980s.⁸ Various types of laser used in treatment such as low-power lasers are He-Ne (632.5 nm), diode lasers (DLs) with various wavelengths 810, 940, and 980 nm, and medium-power lasers as Nd: YAG (1064 nm), CO₂ (10600 nm), Er: YAG (2940 nm), and Er, Cr: YSGG (2780 nm). In these lasers, He-Ne (632.5 nm) and Diode Lasers have analgesic effect, they can have effect alone or in combination with desensitizing agents.

The desensitizing effect of the high power laser is thought to be related to the occlusion or narrowing of the dentinal tubules. These effects of these lasers are related to an increase in surface temperature which can result in the complete closure of dentinal

tubules after recrystallization of the dentinal surface. According to hydrodynamic theory, such occlusion could mediate stimulus transmission from the dental surface to the pulp.⁹ Low-level laser therapy (LLLT) is a sensitizing method to promote biomodulatory effects, minimize pain and reduce inflammatory processes that shows a great promise. This treatment induces alterations within the net of nerve transmission of the dental pulp, instead of altering the exposed dentinal surface, as in most other types of treatment. LLLT has been used for DH since the 1980s. Studies using the gallium aluminum arsenide (GaAlAs) laser showed DH reduction in the range of 60%–98%. The aim of this paper is to review effects and critically discuss most relevant aspects related to LLLT (Low level laser therapy) in minimizing dentinal pain.

LLLT and its Clinical Applications

The application of low-level lasers in medicine was introduced in the 1980s. Since then considerable scientific work including the use of cell cultures, animal models and clinical studies has been undertaken to evaluate its potentially beneficial effects. The application of LLLT has become popular in a variety of clinical applications, including promotion of wound healing and reduction of pain. Low level laser applications in dentistry include the promotion of wound healing in a range of sites, like surgical wounds, extraction sites, recurrent aphthous ulcerations, etc. Applications of LLLT in dental and periodontal treatments represent the subject of many in vivo and in vitro studies, which recommend the use of laser therapy after gingivectomy and gingivoplasty procedures due to its ability to speed up the healing process.

The low-level lasers facilitate fibroblast and keratinocyte motility, collagen synthesis, angiogenesis and growth factors release, thus facilitating the healing process. This therapy has been used in pain management protocols following gingivectomies, and as an adjunct treatment in nonsurgical periodontal procedures.¹³⁻¹⁴ Low-level laser therapy (LLLT) has also been considered as an alternative treatment option for pain in dentinal hypersensitivity. Kimura et al. summarized the current knowledge regarding laser applications for

the treatment of dentin hypersensitivity.¹⁵ Several other studies evaluated the effectiveness of the clinical use of diode lasers for the treatment of dentin hypersensitivity and reported their use as effective in reducing dentinal hypersensitivity.¹⁶⁻²² Three wavelengths (780, 830, and 900 nm), all within the infrared spectrum of (gallium-aluminium-arsenide) GaAlAs diode laser, have been used for the treatment of dentin hypersensitivity.

Different Low Level Lasers for Dentinal Hypersensitivity

Low power output laser therapy was initially used to support wound healing. Subsequently, in the 1980s, the benefit of LLLT delivery systems were used as an anti-inflammatory tool. After that, it was demonstrated that LLLT systems stimulates nerve cells in a clinical environment.²³⁻²⁵

The low level lasers used for the treatment of dentinal hypersensitivity are of two types:

1. (He-Ne) Helium neon laser (632nm)
2. (Ga-Al-As) Gallium aluminium arsenide or diode laser (655nm to 980nm)

(He-Ne) Helium neon laser (632nm)

The first low-level laser introduced was helium-neon (He-Ne), which combined a gaseous mixture to produce a wavelength in the visible light spectrum ($\lambda = 632.8 \text{ nm}$) and low power output (ranging from 5 mW to 30 mW). Since the wavelength produced by He-Ne laser was highly absorbed by soft tissues, its penetration was limited. The first use of this laser for the treatment of dentine hypersensitivity was reported by Senda et al. (1985), then, consecutively by several other investigators.²⁶ Irradiation modes were two types: pulsed (5 Hz only) and continuous wave (CW) mode. The laser tip has to be placed as close as possible to the tooth surface in noncontact mode. The mechanism involved is mostly unknown. Treatment effectiveness rates of He-Ne laser ranges from 5.2%– 100% based on different studies. According to physiological experiments, He-Ne laser irradiation does not affect peripheral A-delta or C-fiber nociceptors, but does affect electric activity (action potential) which in the healthy nerve increased by 33% following a single transcutaneous irradiation.²⁷⁻²⁸

This was found to be a long-lasting effect, inducing an increase in the size of nerve action potential for more than 8 months after cessation of irradiation. He-Ne laser irradiation at 6 mW does not affect the enamel or dentin surface morphologically, but a small fraction of the laser energy is transmitted through enamel or dentin to reach the pulp tissue. With low power-output lasers, there is no danger of causing skin burns or damaging cells.²⁹⁻³¹



Fig 1: He-Ne laser device

Indications: Mild cervical dentin hypersensitivity (score 1) is only one indication (acc. To VAS score). In the cases of moderate or severe dentin hypersensitivity (score 2 or 3), the effects cannot be expected.

Parameters: 6 mW and 5 Hz or continuous wave (CW) mode for 2 to 5 minutes.

Technique: The laser tip has to be placed as close as possible to the tooth surface in non contact mode. The irradiation is applied to the same tooth surface without scanning. The examination of change of dentin hypersensitivity is carried out every 30 minutes until the dentin hypersensitivity decreases.³²⁻³⁸

(GaAlAs) Gallium-Aluminium-Arsenide laser or Diode laser

Diode lasers are usually variants of gallium:aluminum:arsenide (GaAlAs), which emit in the near infrared spectrum (780 nm, 830 nm, and 900 nm; power output from 20 to 100 mW), or indium:gallium:arsenide:phosphorus (In:Ga:As:P) devices, which emit wavelengths in the red spectrum of visible light (600 to 680 nm, power output from 1 to 50 mW). In their early stages of development, GaAs systems were difficult to run for long periods in a CW mode because of the propensity of the chip to overheat. However, by

1979, experiments using a new diode were looking very promising. This new chip, which used wafer-thin crystals of GaAlAs, could produce a variety of wavelengths from 720 to 904 nm, all within the infrared spectrum. It could also generate a continuous wave with no likelihood of overheating. Mainly four wavelengths (780, 790, 830, and 900 nm) of GaAlAs laser have been used for the treatment of dentin hypersensitivity. New diode lasers were developed in the attempt to obtain slightly higher power output and wavelengths that could penetrate soft tissues without damaging them.

Indications: Mild and moderate dentin hypersensitivity (score 1 and 2) are indications.

Parameters: The laser devices of 3 W with the wavelength of 810 nm (Osada, Tokyo, Japan) and 0.5- 20 W with the wavelength of 805 nm (Panasonic, Osaka, Japan) are sold now. The 3 W power output device is a CW laser device, and the 0.5-20 W power output device is CW or pulsed mode, with a pulse width range from 0.003 to 0.2 s. As the guide beam, a semiconductor laser of the 635 nm wavelength and 1 mW power is used.

Technique: In order to prevent thermal damage to the dental pulp by lasers, the laser tip is kept more than 5 cm from the tooth surface. Furthermore, when the patient feels pain, the laser tip has to be scanned quickly over the tooth surface. Sometimes, the air spray is also used for preventing a temperature rise on the surface during laser irradiation.



Fig 2: GaAlAs laser (Diode Laser)

Mechanism of pain reduction by Low Level Lasers

The desensitizing mechanism obtained with LLLT is as yet elusive. It is believed that low-level

therapy lasers stimulate nerve cells, interfering with the polarity of cell membranes by increasing the amplitude of the action potentials of cellular membranes, thus blocking the transmission of pain stimuli in hypersensitive dentin. It seems that the low output lasers mediate analgesic effects due to depressed nerve transmission.³⁹⁻⁴⁰

Also according to the recent literatures, the low absorbed energy by the dentin surface (via its mineral such as phosphate and carbonate) leads to heat accumulation, which gradually increases the surface temperature (Ying, Gao et al., 2013).⁴¹ This results in denaturation and modification of organic matrix layer with an amorphous form and hence DTs sealing (Marchesan, Brugnera-Junior et al., 2008).⁴²

DISCUSSION

Low-power laser therapy is an appropriate treatment strategy to promote biomodulatory effects, minimize pain and reduce inflammatory processes. Its use has been widely accepted and approved due to satisfactory results reported in the literatures. Although the desensitization mechanisms produced by HLLT i.e high-power lasers, such as the carbon dioxide, Nd:YAG, Er:YAG and Er,Cr:YSGG lasers have been widely discussed, are related to an increase in surface temperature which can result in the complete closure of dentinal tubules and image analysis also detected obliteration of tubules after laser irradiation.⁴³ Special concern when deciding to use laser irradiation in the management of hypersensitive dentin are the precise irradiation parameters required for therapeutic laser effects and standardization of treatment protocols. Defined and appropriate parameters are extremely important to produce the effects desired. Different outcomes may be produced depending on the parameters employed.⁴⁴⁻⁴⁷

Low-level laser therapy is more effective than high energy irradiation. LLLT have no adverse effects have been reported with the use of such therapy to treat hypersensitive teeth. In addition to the biological effect of increasing the potential of action of the pulp tissue, authors emphasize that low energy wavelengths produced by LLLT are safer to the pulp because they stimulate circulation

and cellular activity. However, LLLT does not produce any changes in mineralized tooth substrate. Due to the individual differences observed in response to the laser therapy, additional sessions of LLLT may be necessary in order to obtain a positive result. Since low-level energy irradiation is mostly related to biostimulation and analgesia, it seems obvious that such effects are mainly temporary. The role of the “soft laser” as a therapeutic tool is a contentious issue. According to WilderSmith, clinical trials demonstrated no advantage in replacing conventional treatment of hypersensitive dentin with low-level laser therapy, despite its positive effect on patient attitude toward treatment. Contrasting results, however, have supported LLLT. WilderSmith reported that more positive effects were observed right after the first low-level laser irradiation, whereas cumulative effects and a gradual improvement from visit to visit should be expected.⁴⁸⁻⁵¹

In double-blinded studies this may be particularly true, when information given at the beginning of the study may influence patient’s perception toward the treatment. Pulpal effects of the laser devices previously discussed have been investigated in various studies. The GaAlAs laser at a wavelength of 780 nm, and a power output of 30 mW for 3 min caused no thermal or other damage to pulp tissues in monkeys.⁵² According to an in vitro thermometric study, GaAlAs laser irradiation at the parameters of 30 mW (CW) at 780 nm wavelength, 60 mW (CW) at 830 nm wavelength, and 10 W (pulsed) at 900 nm wavelength do not cause significant intrapulpal temperature rises.⁵³

The LLLT assisted treatment of dentine hypersensitivity is a good method to solve immediate and long term pain, compared to conventional desensitizing topical agents and the high level laser treatment. It is although more expensive but leads to rapid results with less

application time and more quickly for the patient. In most of the articles, fluoride gel or desensitizing substances used in combination with LLLT can potentiate effects. The same line of reasoning is considered valid for the association with desensitizing pastes. New substances as cyanoacrylate, glutaraldehyde and potassium binoxalate are spreading for the properties to stimulate laser beneficial effects and they can be used alone as preventative measures in patients with mild hypersensitivity. However, the effectiveness of these treatments has clashed sometimes with the existence of a placebo effect. In the majority of studies, patients have a decrease in VAS from baseline both immediately and over time, till six months after treatment. The diode laser appears to be the most widely used in everyday practice by dental hygienists and dentists. Studies are clarifying the follow-up results within the interference of the placebo effect. However, in vitro studies confirm a real effectiveness of these lasers. Thanks to the SEM analysis, the percentage of occlusion appears to be complete and the diameter of dentinal tubules reduced.

CONCLUSION

LLLT, due to its reduced side effects can be considered as an effective treatment for the dentinal hypersensitivity. All the parameters for the treatment are already set up by the manufacturer and supplied with specific protocols for each treatment. In consideration with all the literatures, it can be said that low level laser is an innovative and faster treatment both in terms of therapy time and results, with minimal side effects and greater comfort for patients, which appear more satisfied with traditional methods. Although it would seem that the LLLT effectively reduces pain symptoms, further studies and more suitable follow-ups are necessary.

REFERENCES

1. C. R. Irwin and P. McCusker, “Prevalence of dentine hypersensitivity in a general dental population,” *Journal of the Irish Dental Association*, vol. 43, no. 1, pp. 7–9, 1997.
2. Flynn, J., Galloway, R., and Orchardson, R. (1985). The incidence of “hypersensitive” teeth in the West of Scotland. *J. Dent.* 13, 230–236.

3. Orchardson, R., and Collins, W.J.N. (1987). Clinical features of hypersensitive teeth. *Br. Dent. J.* 162, 253–256.
4. Dowell, P., Addy, M., and Dummer, P. (1985). Dentine hypersensitivity: aetiology, differential diagnosis and management. *Br. Dent. J.* 158, 92–96.
5. M. Brannstrom, “A hydrodynamic mechanism in the transmission of pain-producing stimuli through dentine,” in *Sensory Mechanism in Dentine*, D. J. Anderson, Ed., pp. 73–79, Pergamon, Oxford, UK, 1963.
6. T. C. C. G. P. Ladalardo, A. Pinheiro, R. A. D. C. Campos et al., “Laser therapy in the treatment of dentine hypersensitivity,” *Brazilian Dental Journal*, vol. 15, no. 2, pp. 144–150, 2004.
7. Goodis HE, White JM, Marshall Jr GW, Yee K, Fuller N, Gee L, Marshall SJ. Effects of Nd: and Ho: Yttrium-Aluminum-Garnet lasers on human dentine fluid flow and dental pulp chamber temperature in vitro. *Arch Oral Biol* 1997;42:845-854.
8. Porto IC, Andrade AK, Montes MA. Diagnosis and treatment of dentinal hypersensitivity. *J Oral Sci* 2009;51:323-32.
9. Addy M. Etiology and clinical implications of dentine hypersensitivity. *Dent Clin North Am* 1990;34:503-14.
10. Suri I, Singh P, Shakir QJ, Shetty A, Bapat R, Thakur R, et al. A comparative evaluation to assess the efficacy of 5% sodium fluoride varnish and diode laser and their combined application in the treatment of dentin hypersensitivity. *J Indian Soc Periodontol* 2016;20:307-14.
11. Marsilio AL, Rodrigues JR, Borges AB. Effect of the clinical application of the GaAlAs laser in the treatment of dentine hypersensitivity. *J Clin Laser Med Surg* 2003;21:291-6.
12. Neiburger EJ. The effect of low-power lasers on intraoral wound healing. *N Y State Dent J* 1995;61:40-3.
13. Ozcelik O, Cenk Haytac M, Kunin A, Seydaoglu G. Improved wound healing by low-level laser irradiation after gingivectomy operations: A controlled clinical pilot study. *J Clin Periodontol* 2008;35:250-4.
14. Mărtu S, Amalinei C, Tatarciuc M, Rotaru M, Potârniche O, Liliac L, et al. Healing process and laser therapy in the superficial periodontium: A histological study. *Rom J Morphol Embryol* 2012;53:111-6.
15. Kimura Y, Wilder-Smith P, Yonaga K, Matsumoto K (2000) Treatment of dentine hypersensitivity by lasers: a review. *J Clin Periodontol* 27:715–721.
16. Tengrungsun T, Sangkla W (2008) Comparative study in desensitizing efficacy using the GaAlAs laser and dentin bonding agent. *J Dent* 36:392–395
17. Ladalardo TC, Pinheiro A, Campos RA et al (2004) Laser therapy in the treatment of dentine hypersensitivity. *Braz Dent J* 15:144–150
18. Corona SA, Nascimento TN, Catirse AB, Lizarelli RF, Dinelli W, Palma-Dibb RG (2003) Clinical evaluation of low-level laser therapy and fluoride varnish for treating cervical dentinal hypersensitivity. *J Oral Rehabil* 30:1183–1189
19. Marsilio AL, Rodrigues JR, Borges AB (2003) Effect of the clinical application of the GaAlAs laser in the treatment of dentine hypersensitivity. *J Clin Laser Med Surg* 21:291–296
20. Dilsiz A, Canakci V, Ozdemir A, Kaya Y (2009) Clinical evaluation of Nd:YAG and 685-nm diode laser therapy for desensitization of teeth with gingival recession. *Photomed Laser Surg* 27:843–848
21. Sicilia A, Cuesta-Frechoso S, Suárez A, Angulo J, Pordomingo A, De Juan P (2009) Immediate efficacy of diode laser application in the treatment of dentine hypersensitivity in periodontal maintenance patients: a randomized clinical trial. *J Clin Periodontol* 36:650–660
22. Vieira AH, Passos VF, de Assis JS, Mendonça JS, Santiago SL (2009) Clinical evaluation of a 3% potassium oxalate gel and a GaAlAs laser for the treatment of dentinal hypersensitivity. *Photomed Laser Surg* 27:807–812.
23. Kimura Y, Takebayashi H, Iwase T, Nara Y, Morioka T. Effect of helium-neon laser irradiation on wound healing in rats. *Surg Med Lasers* 1991;4:14-16.
24. Kimura Y, Takebayashi H, Iwase T, Morioka T. The enhancement of transforming growth factor-β activity by helium-neon laser on wound healing in rats. *Lasers Life Sci* 1993;5:209-217.
25. Kimura Y, Iwase T, Morioka T, Wilder-Smith P. Possible platelet-derived growth factor involvement on helium-neon laser stimulated wound healing in rats. *Lasers Life Sci* 1997;7:267-284.
26. Gomi, A., Kamiya, K., Yamashita, H., Ban, Y., Senda, A., Hara, G., Yamaguchi, M., Narita, T. & Hasegawa, J. (1986) A clinical study on “Soft Laser 632”, a HeNe low energy medical laser. *Aichi-Gakuin Journal of Dental Science* 24, 390–399.

27. Rochkind S, Nissan M, Razon N, Schwartz M, Bartal A. Electrophysiological effect of HeNe laser on normal and injured sciatic nerve in the rat. *Acta Neurochir (Wien)* 1986;83:125-130.
28. Rochkind S, Nissan M, Barr-Nea L, Razon N, Schwartz M, Bartal A. Response of peripheral nerve to He-Ne laser: experimental studies. *Lasers Surg Med* 1987;7:441-443.
29. Kimura Y, Wilder-Smith P, Yonaga K, Matsumoto K. Treatment of dentine hypersensitivity by lasers: a review. *J Clin Periodontol.* 2000 Oct; 27(10):715-21.
30. Matsumoto K, Kimura Y. Laser Therapy of Dentin Hypersensitivity. *J Oral Laser Application* 2007; 7: 7-25.
31. Ladalardo TC, Pinheiro A, Campos RA, Brugnera Júnior A, Zanin F, Albernaz PL, Laser therapy in the treatment of dentine hypersensitivity. *Braz. Dent. J.* 2004; 15 (2): 144-150.
32. Senda A, Gomi A, Tani T, Yoshino H, Hara G, Yamaguchi M, Matsumoto T, Narita T, Hasegawa J. A clinical study on "Soft Laser 632", a He-Ne low energy medical laser. *Aichi-Gakuin J Dent Sci* 1985;23:773-780.
33. Matsumoto K, Nakamura G, Tomonari H. Study on the treatment of hypersensitive dentine by He-Ne laser irradiation. *Japan J Conserv Dent* 1986;29:312-317.
34. Gomi A, Kamiya K, Yamashita H, Ban Y, Senda A, Hara G, Yamaguchi M, Narita T, Hasegawa J. A clinical study on "Soft Laser 632", a He-Ne low energy medical laser. *Aichi-Gakuin J Dent Sci* 1986;24:390-399.
35. Kanamura N, Saitoh H, Uematsu N, Morimoto I, Takeda G, Hori N. Pain-relieving effects of soft laser (He-Ne) irradiation on various oral lesions. *Japan J Conserv Dent* 1986;29:1548-1555.
36. Wilder-Smith P. The soft laser: therapeutic tool or popular placebo? *Oral Surg Oral Med Oral Pathol* 1988;66:654-658.
37. Matsumoto K, Nishihama R, Onodera A, Wakabayashi H. Study on treatment of hypersensitive dentine by He-Ne laser. *J Showa Univ Dent Soc* 1988;8:180-184.
38. Mezawa S, Shiono M, Sato K, Mikami T, Hayashi M, Maeda K, Ogawa M, Saito T. The effect of low-power laser irradiation on hypersensitive dentin: differing effect according to the irradiated area. *J Japan Soc Laser Dent* 1992;3:87-91.
39. Gillam DG, Mordan NJ, Newman HN. The dentin disc surface: a plausible model for dentin physiology and dentin sensitivity evaluation. *Adv Dent Res* 1997;11:487-501.I
40. Gerschman JA, Ruben J, Gerbart-Eaglemont J. Low level laser therapy for dentinal tooth hypersensitivity. *Am Dent J* 1994; 39:353-357.
41. Ying, L., Gao, J., Gao, Y., Shuaimei, X., Zhan, X. and Wu, B. (2013) "In vitro study of dentin hypersensitivity treated by 980-nm diode laser." *Journal of lasers in medical sciences* 4(3): 111.
42. Marchesan, M.A., Brugnera-Junior, A., Souza-Gabriel, A. E., Correa-Silva, S.R. and Sousa-Neto, M.D. (2008) "Ultrastructural analysis of root canal dentine irradiated with 980-nm diode laser energy at different parameters." *Photomedicine and laser surgery* 26(3): 235-240.
43. Aranha A, Eduardo C. Effects of Er:YAG and Er,Cr:YSGG lasers on dentine hypersensitivity. Short-term clinical evaluation. *Lasers Med Sci* 2012; 27:813-818.
44. Walsh LJ. The current status of low level laser therapy in dentistry. Part 2. Hard tissue applications. *Austr Dent J* 1997; 42:302-306
45. Goodis HE, White JM, Marshall Jr GW, Yee K, Fuller N, Gee L, Marshall SJ. Effects of Nd: and Ho: Yttrium-Aluminum-Garnet lasers on human dentine fluid flow and dental pulp chamber temperature in vitro. *Arch Oral Biol* 1997;42:845-854
46. Liu HC, Lin CP, Lan WH. Sealing depth of Nd:YAG laser on human dentinal tubules. *J Endod* 1997;23:691-693
47. Pashley EL, Horner JÁ, Liu M, Kim S, Pashley DH. Effects of CO2 laser energy on dentin permeability. *J Endod* 1992;18:257-
48. Wilder-Smith P. The soft laser: therapeutic tool or popular placebo? *Oral Surg Oral Med Oral Pathol* 1988;66:654-658.
49. Wakabayashi H, Matsumoto K. Treatment of dentine hypersensitivity by GaAlAs soft laser irradiation [abstract 554]. *J Dent Res* 1988;67:182.
50. Gelskey SC, White JM, Pruthi VK. The effectiveness of the Nd:YAG laser in the treatment of dental hypersensitivity. *J Can Dent Assoc* 1993;59:377-386.
51. Marsilio AL, Rodrigues JJ, Borges AB. Effect of the clinical application of the GaAlAs laser in the treatment of dentine hypersensitivity. *J Clin Laser Med Surg* 2003;21:291-296.
52. Matsumoto K, Wakabayashi H, Funato A, Shirasuka T. Histopathological findings of dental pulp irradiated by GaAlAs laser diode. *Japan J Conserv Dent* 1985;28:1361-1365

53. Arrastia AMA, Machida T, Wilder-Smith P, Matsumoto K. Comparative study of the thermal effects of four semiconductor lasers on the enamel

and pulp chamber of a human tooth. *Lasers Surg Med* 1994;15:382-389.